

Absorption Spectrometer



UV-Vis Array Spectrometer
UV-Vis Light Source
Absorbance Cell
Optical Fibers

طيف سنج جذبی

Abs - Ar - 2015

Specifications of Absorption Array Spectrometer

Optical Specifications

Wavelength range	200-800 nm(190-1100nm)
Straylight	< 0.03%
Resolution	< + 1.0 nm
Wavelength Accuracy	< ± 0.1 nm
Wavelength Reproducibility	< ± 0.02 nm for 10 consecutive scans (NIST 2034)
Photometric Accuracy	< ± 0.01 A at 440.0 465.0 590.0 635.0 nm(NIST 930e)
Photometric noise	<0.01 A 80 scan at 0 A, 500 nm
Photometric stability	<0.01 A/min at 0 A, 500 nm
Baseline flatness	<0.01A 0.5 second blank, 0.5 second scan at 0 A, 500 nm
Scan time	2-1000 ms
Typical Scan Time	0.01s
Light Source	Deuterium + Halogen /Tungsten lamp
Detector	2000 Multichannel

Physical Specifications

Communication	USB 2
Software	Windows XP , 7, 8
Dimensions (Spectrometer)	13 x 22 x 28 cm
Dimensions (Light Source)	8 x 22 x 28 cm
Weight	8 kg
Power	220V AC 50/60 HZ

Tell: +98 21 88357692

Fax: +98 21 88357803

Address: No.115, Dr. Fathi shaghghi St, Yosef abad, Tehran, Iran

Teifsanje Tadjhiz Pishrafteh Co. Ltd.

Email :info@teifsanje.ir

Website :www.teifsanje.ir

